

DePuy Synthes CONDUIT™ ZERO-P™ VA Secured Spacer System

Patient Implant Card

INSTRUCTIONS

Please print and provide the Patient Implant Card to your DePuy Synthes CONDUIT ZERO-P VA patient. Please fill out the front page of the Implant Card.

EXPLANATION of the IMPLANT CARD:

The Patient Implant Card is intended for US patients who receive DePuy Synthes CONDUIT ZERO-P VA implants. The Implant Card provides MRI safety status of the DePuy Synthes CONDUIT ZERO-P VA implants and URL/contact information where full MRI safety information can be found.

It is recommended that patients carry the Implant Card at all times to identify themselves as individuals who have a DePuy Synthes CONDUIT ZERO-P VA implant. This Implant Card is useful for situations in which patients may need MRI scanning or must go through a metal detection device.

For access to the e-IFU, visit www.avalign.com/ifu.

RD_IFU-23001B-IC, Rev. 1
01 APR 2026

DePuy Synthes

USA
Raynham, MA 02767-0350
325 Paramount Drive
DePuy Synthes Sales, Inc.

Avalign Technologies, Inc.
8727 Clinton Park Drive
Fort Wayne, IN 46825 USA
1-877-289-1096
product:questions@avalign.com
www.avalign.com/ifu

Use, which can be obtained at www.avalign.com/ifu.
Magnetic Resonance (MR) Compatibility section of the Instructions for
device malfunction. Full MRI safety information is available in the
conditions. Scanning under different conditions may result in injury or
device and can safely undergo an MR exam only under very specific
This person is implanted with an MR Conditional CONDUIT ZERO-P VA

MR CONDITIONAL

MD

CONDUIT™ ZERO-P™ VA Secured Spacer

System

PATIENT IMPLANT CARD

Patient Name

Healthcare Centre or Doctor

Surgery Date

Show this card to your healthcare provider and inform MRI site
personnel that you have an **MR Conditional** medical device during MR
screening, prior to the MR exam.